





TECHNICAL APPENDIX 14.2: CALIBRATION CERTIFICATES





| | | |
|-----------|--|------------|
| 1. | CALIBRATION CERTIFICATE NL52 01265412 | 1-2 |
| 2. | CALIBRATION CERTIFICATE NL52 00175536 | 2-4 |
| 3. | CALIBRATION CERTIFICATE NC74 34178103 | 3-6 |

1. CALIBRATION CERTIFICATE NL52 01265412

| | | | | | |
|--|--|---|---|---|---|
|  | | CERTIFICATE OF CALIBRATION | |  |  0653 |
| Date of Issue: 21 April 2022 Calibrated at & Certificate issued by: ANV Measurement Systems Beaufort Court 17 Roebuck Way Milton Keynes MK5 8HL Telephone 01908 642846 Fax 01908 642814 E-Mail: info@noise-and-vibration.co.uk Web: www.noise-and-vibration.co.uk <small>Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems</small> | | | Certificate Number: UCRT22/1552 | | |
| | | | Page 1 of 2 Pages | | |
| | | | Approved Signatory | | |
| | | |  | | |
| | | | K. Mistry | | |
| Customer | Wood Group UK Ltd St Vincent Plaza St Vincent Street Glasgow G2 5LD | | | | |
| Order No. | 26010406 | | | | |
| Description | Sound Level Meter / Pre-amp / Microphone / Associated Calibrator | | | | |
| Identification | <i>Manufacturer</i> | <i>Instrument</i> | <i>Type</i> | <i>Serial No. / Version</i> | |
| | Rion | Sound Level Meter | NL-52 | 01265412 | |
| | Rion | Firmware | | 2.0 | |
| | Rion | Pre Amplifier | NH-25 | 65414 | |
| | Rion | Microphone | UC-59 | 10633 | |
| | Brüel & Kjær | Calibrator | 4231 | 2052327 | |
| | | Calibrator adaptor type if applicable | | UC 0210 | |
| Performance Class | 1 | | | | |
| Test Procedure | TP 10. SLM 61672-3:2013 <i>Procedures from IEC 61672-3:2013 were used to perform the periodic tests.</i> | | | | |
| Type Approved to IEC 61672-1:2013 | Yes | | | | |
| | <i>If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013</i> | | | | |
| Date Received | 19 April 2022 | | ANV Job No. | UKAS22/04276 | |
| Date Calibrated | 21 April 2022 | | | | |
| <p>The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.</p> | | | | | |
| Previous Certificate | <i>Dated</i> | <i>Certificate No.</i> | <i>Laboratory</i> | | |
| | 22 May 2020 | UCRT20/1446 | 0653 | | |
| <p>This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.</p> | | | | | |

| CERTIFICATE OF CALIBRATION | | | | Certificate Number | |
|---|------------------|-------------------------|--|---------------------------|------|
| UKAS Accredited Calibration Laboratory No. 0653 | | | | UCRT22/1552 | |
| Page 2 of 2 Pages | | | | | |
| Sound Level Meter Instruction manual and data used to adjust the sound levels indicated. | | | | | |
| SLM instruction manual title NL-52/NL-42 Description for IEC 61672-1 | | | | | |
| SLM instruction manual ref / issue No. 56034 21-03 Source Rion | | | | | |
| Date provided or internet download date 19 March 2021 | | | | | |
| Uncertainties provided | Case Corrections | Wind Shield Corrections | Mic Pressure to Free Field Corrections | | |
| Yes | Yes | Yes | | | |
| Total expanded uncertainties within the requirements of IEC 61672-1:2013 | | | | | YES |
| Specified or equivalent Calibrator Equivalent | | | | | |
| Customer or Lab Calibrator Customers Calibrator | | | | | |
| Calibrator adaptor type if applicable UC 0210 | | | | | |
| Calibrator cal. date 20 April 2022 | | | | | |
| Calibrator cert. number UCRT22/1540 | | | | | |
| Calibrator cal cert issued by Lab 0653 | | | | | |
| Calibrator SPL @ STP 94.11 dB Calibration reference sound pressure level | | | | | |
| Calibrator frequency 999.79 Hz Calibration check frequency | | | | | |
| Reference level range Single dB | | | | | |
| Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15 | | | | | |
| Note - The Extension Cable was used between the SLM and the pre-amp for this calibration. | | | | | |
| Environmental conditions during tests | | | | | |
| | Start | End | | | |
| Temperature | 24.50 | 24.30 | ± 0.30 °C | | |
| Humidity | 40.6 | 42.2 | ± 3.00 %RH | | |
| Ambient Pressure | 100.24 | 100.20 | ± 0.03 kPa | | |
| Indication at the Calibration Check Frequency | | | | | |
| Initial indicated level | 94.2 | dB | Adjusted indicated level | 94.1 | dB |
| Uncertainty of calibrator used for Indication at the Calibration Check Frequency ± | | | | | 0.10 |
| Self Generated Noise | | | | | |
| Microphone installed - Less Than 17.9 dB A Weighting | | | | | |
| Microphone replaced with electrical input device - UR = Under Range indicated | | | | | |
| Weighting | A | C | Z | | |
| | 11.7 | 15.5 | 20.8 | dB | UR |
| Self Generated Noise reported for information only and not used to assess conformance to a requirement | | | | | |
| The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements. | | | | | |
| <u>Additional Comments</u> The results on this certificate only relate to the items calibrated as identified above. | | | | | |
| None | | | | | |
| END | | | | | |
| Calibrated by: B. Giles | | | | R 3 | |

2. CALIBRATION CERTIFICATE NL52 00175536

| | | | |
|--|--|---|---|
|  | <p>CERTIFICATE OF CALIBRATION</p> |  |  0653 |
| <p>Date of Issue: 25 May 2022</p> <p>Calibrated at & Certificate issued by: ANV Measurement Systems Beaufort Court 17 Roebuck Way Milton Keynes MK5 8HL Telephone 01908 642846 Fax 01908 642814 E-Mail: info@noise-and-vibration.co.uk Web: www.noise-and-vibration.co.uk <small>Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems</small></p> | | <p>Certificate Number: UCRT22/1695</p> | |
| | | <p>Page 1 of 2 Pages</p> | |
| | | <p>Approved Signatory</p>  | |
| | | <p>K. Mistry</p> | |
| Customer | Wood Group St. Vincent Plaza (Floor 2) 319 St. Vincent Street Glasgow G2 5LP | | |
| Order No. | 26010406 | | |
| Description | Sound Level Meter / Pre-amp / Microphone / Associated Calibrator | | |
| Identification | <i>Manufacturer</i> | <i>Instrument</i> | <i>Type</i> |
| | Rion | Sound Level Meter | NL-52 |
| | Rion | Firmware | 2.0 |
| | Rion | Pre Amplifier | NH-25 |
| | Rion | Microphone | UC-59 |
| | Rion | Calibrator | NC-74 |
| | | Calibrator adaptor type if applicable | NC-74-002 |
| Performance Class | 1 | | |
| Test Procedure | TP 10. SLM 61672-3:2013 <i>Procedures from IEC 61672-3:2013 were used to perform the periodic tests.</i> | | |
| Type Approved to IEC 61672-1:2013 | Yes | | |
| | <i>If YES above there is public evidence that the SLM has successfully completed the applicable pattern evaluation tests of IEC 61672-2:2013</i> | | |
| Date Received | 23 May 2022 | ANV Job No. | UKAS22/05346 |
| Date Calibrated | 25 May 2022 | | |
| <p>The sound level meter submitted for testing has successfully completed the periodic tests of IEC 61672-3:2013, for the environmental conditions under which the tests were performed. As evidence was publicly available, from an independent testing organisation responsible for approving the results of pattern-evaluation tests performed in accordance with IEC 61672-2:2013, to demonstrate that the model of sound level meter fully conformed to the class 1 specifications in IEC 61672-1:2013, the sound level meter submitted for testing conforms to the class 1 specifications of IEC 61672-1:2013.</p> | | | |
| Previous Certificate | <i>Dated</i> | <i>Certificate No.</i> | <i>Laboratory</i> |
| | 26 May 2020 | UCRT20/1449 | 0653 |
| <p>This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.</p> | | | |

| | |
|-----------------------------------|--|
| CERTIFICATE OF CALIBRATION | Certificate Number UCRT22/1695 |
| | Page 2 of 2 Pages |

UKAS Accredited Calibration Laboratory No. 0653

Sound Level Meter Instruction manual and data used to adjust the sound levels indicated.

| | | | |
|--|---|-------------------------|--|
| SLM instruction manual title | NL-52/NL-42 Description for IEC 61672-1 | | |
| SLM instruction manual ref / issue | No. 56034 21-03 | Source | Rion |
| Date provided or internet download date | 19 March 2021 | | |
| | Case Corrections | Wind Shield Corrections | Mic Pressure to Free Field Corrections |
| Uncertainties provided | Yes | Yes | Yes |
| Total expanded uncertainties within the requirements of IEC 61672-1:2013 | | | |
| | YES | | |
| Specified or equivalent Calibrator | Specified | | |
| Customer or Lab Calibrator | Customers Calibrator | | |
| Calibrator adaptor type if applicable | NC-74-002 | | |
| Calibrator cal. date | 24 May 2022 | | |
| Calibrator cert. number | UCRT22/1682 | | |
| Calibrator cal cert issued by Lab | 0653 | | |
| Calibrator SPL @ STP | 94.02 | dB | Calibration reference sound pressure level |
| Calibrator frequency | 1001.97 | Hz | Calibration check frequency |
| Reference level range | Single dB | | |

Accessories used or corrected for during calibration - Extension Cable & Wind Shield WS-15

Note - The Extension Cable was used between the SLM and the pre-amp for this calibration.

| | | | | |
|---------------------------------------|------------------|--------|--------|------------|
| Environmental conditions during tests | | Start | End | |
| | Temperature | 24.65 | 24.37 | ± 0.30 °C |
| | Humidity | 48.5 | 48.7 | ± 3.00 %RH |
| | Ambient Pressure | 100.05 | 100.05 | ± 0.03 kPa |

| | | | |
|--|------|---------|--------------------------|
| Indication at the Calibration Check Frequency | | | |
| Initial indicated level | 94.3 | dB | Adjusted indicated level |
| | | | 94.0 dB |
| Uncertainty of calibrator used for indication at the Calibration Check Frequency ± | | 0.10 dB | |
| Self Generated Noise | | | |

| | | | | |
|--|------------|----------------------------|------------|-------------|
| Microphone installed - | Less Than | 19.1 | dB | A Weighting |
| Microphone replaced with electrical input device - | | UR = Under Range indicated | | |
| Weighting | A | C | Z | |
| | 13.1 dB UR | 16.9 dB UR | 23.1 dB UR | |

Self Generated Noise reported for information only and not used to assess conformance to a requirement

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor $k=2$, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.



Additional Comments The results on this certificate only relate to the items calibrated as identified above.

None

..... END

Calibrated by: B. Bogdan R 2

3. CALIBRATION CERTIFICATE NC74 34178103

| | | | | |
|---|--|---|---|-------------------------------|
|  | CERTIFICATE OF CALIBRATION |  |  0653 | |
| Date of Issue: 24 May 2022 Calibrated at & Certificate issued by: ANV Measurement Systems Beaufort Court 17 Roebuck Way Milton Keynes MK5 8HL Telephone 01908 642846 Fax 01908 642814 E-Mail: info@noise-and-vibration.co.uk Web: www.noise-and-vibration.co.uk | Certificate Number: UCRT22/1682 | Page 1 of 2 Pages | | |
| Acoustics Noise and Vibration Ltd trading as ANV Measurement Systems | | Approved Signatory  K. Mistry | | |
| Customer | Wood Group St. Vincent Plaza (Floor 2) 319 St. Vincent Street Glasgow G2 5LP | | | |
| Order No. | 26010406 | | | |
| Test Procedure | Procedure TP 1 Calibration of Sound Calibrators | | | |
| Description | Acoustic Calibrator | | | |
| Identification | <i>Manufacturer</i> Rion | <i>Instrument</i> Calibrator | <i>Model</i> NC-74 | <i>Serial No.</i> 34178103 |
| <p>The calibrator has been tested as specified in Annex B of IEC 60942:2003. As public evidence was available from a testing organisation (PTB) responsible for approving the results of pattern evaluation tests, to demonstrate that the model of sound calibrator fully conformed to the requirements for pattern evaluation described in Annex A of IEC 60942:2003, the sound calibrator tested is considered to conform to all the class 1 requirements of IEC 60942:2003.</p> | | | | |
| ANV Job No. | UKAS22/05346 | | | |
| Date Received | 23 May 2022 | | | |
| Date Calibrated | 24 May 2022 | | | |
| Previous Certificate | <i>Dated</i> 22 May 2020 | <i>Certificate No.</i> UCRT20/1440 | <i>Laboratory</i> 0653 | |
| <p>This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.</p> | | | | |

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. 0653

Certificate Number

UCRT22/1682

Page 2 of 2 Pages

Measurements

The sound pressure level generated by the calibrator in its WS2 configuration was measured five times by the Insert Voltage Method using a microphone as detailed below. The mean of the results obtained is shown below. It is corrected to the standard atmospheric pressure of 101.3 kPa (1013 mBar) using original manufacturers information.

| | | |
|-----------------|---------------------|-------------|
| Test Microphone | <i>Manufacturer</i> | <i>Type</i> |
| | Brüel & Kjær | 4134 |

Results

The level of the calibrator output under the conditions outlined above was

94.02 ± 0.10 dB rel 20 µPa

Functional Tests and Observations

| | |
|---|--------------------------|
| The frequency of the sound produced was | 1001.97 ± 0.12 Hz |
| The total distortion was | 1.61 ± 0.11 % Distortion |

During the measurements environmental conditions were

| | | | | |
|---------------------|------|----|------|-----|
| Temperature | 23 | to | 24 | °C |
| Relative Humidity | 44 | to | 51 | % |
| Barometric Pressure | 99.4 | to | 99.5 | kPa |

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a coverage probability of approximately 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

The uncertainties refer to the measured values only with no account being taken of the ability of the instrument to maintain its calibration.

A small correction factor may need to be applied to the sound pressure level quoted above if the device is used to calibrate a sound level meter which is fitted with a free-field response microphone. See manufacturers handbook for details.

END

Note:

| | |
|---|--------|
| Calibrator adjusted prior to calibration? | NO |
| Initial Level | N/A dB |
| Initial Frequency | N/A Hz |

Additional Comments The results on this certificate only relate to the items calibrated as identified above.

None

Calibrated by: B. Bogdan

R 2